



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

AMERICAN INSTRUMENT CORPORATION

Crossroads Centre  
702 Rose Drive  
Hartland, WI 53029  
Jeff Quinn Phone: 262 367 4409

CALIBRATION

Valid To: December 31, 2024

Certificate Number: 1354.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 5</sup>:

I. Chemical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
pH Meters & Sensors <sup>3</sup>	(4, 7, 10) pH	0.026 pH	MV & temperature, calibration standard buffers

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 7</sup> (±)	Comments
DC Voltage <sup>3</sup> – Generate	(0 to 100) mV (0 to 1) V (0 to 10) V	0.0070 mV 0.075 mV 0.50 mV	Fluke 525B
DC Voltage <sup>3</sup> – Measure	(0 to 75) mV (0 to 10) V	0.0074 mV 0.34 mV	Fluke 8845A
DC Current <sup>3</sup> – Generate	(0 to 100) mA	1.3 µA	Fluke 7526A

Parameter/Equipment	Range	CMC <sup>2,7</sup> ( $\pm$ )	Comments
DC Current <sup>3</sup> – Measure	(0 to 50) mA	0.000 58 mA	Fluke 7526A
Resistance <sup>3</sup> – Generate	(5 to 400) $\Omega$ (5 to 4) k $\Omega$	0.12 $\Omega$ 0.37 $\Omega$	Fluke 525B
Resistance <sup>3</sup> – Measure	(1 to 100) $\Omega$ 100 $\Omega$ to 1 M $\Omega$	(0.0048 + 0.012 %) $\Omega$ (0.11 + 0.013 %) $\Omega$	Fluke 8845A
Electrical Calibration of RTD Indicators <sup>3</sup> – Recorders, Controllers & Calibrators			
Pt 385, 100 $\Omega$	(-200 to 800) °C	0.15 °C	Fluke 525B
Pt 385, 1000 $\Omega$	(-200 to 630) °C	0.14 °C	
Electrical Calibration of Thermocouple Indicators <sup>3</sup> – Recorders, Controllers, & Calibrators			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.97 °C 0.78 °C 0.66 °C 0.66 °C	Fluke 525B
Type C	(0 to 150) °C (150 to 650) °C (650 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C	0.35 °C 0.31 °C 0.37 °C 0.58 °C 0.97 °C	
Type E	(-270 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.67 °C 0.41 °C 0.38 °C 0.39 °C 0.42 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Indicators <sup>3</sup> – Recorders, Controllers, & Calibrators (cont)			
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.32 °C 0.25 °C 0.21 °C 0.22 °C 0.29 °C	Fluke 525B
Type K	(-270 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.40 °C 0.27 °C 0.21 °C 0.32 °C 0.48 °C	
Type N	(-270 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.49 °C 0.30 °C 0.25 °C 0.24 °C 0.34 °C	
Type R	(0 to 250) °C (250 to 600) °C (600 to 1000) °C (1000 to 1750) °C	0.67 °C 0.52 °C 0.48 °C 0.54 °C	
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1750) °C	0.65 °C 0.53 °C 0.51 °C 0.60 °C	
Type T	(-270 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.74 °C 0.30 °C 0.20 °C 0.19 °C	

III. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Indirect Verification of Rockwell Hardness Testers & Rockwell Superficial Hardness <sup>3</sup>	HRA: Low Middle High	0.32 HRA 0.21 HRA 0.18 HRA	Indirect verification method per ASTM E18-22
	HRBW: Low Middle High	0.44 HRBW 0.36 HRBW 0.40 HRBW	
	HRC: Low Middle High	0.44 HRC 0.35 HRC 0.33 HRC	
	HREW: Low Middle High	0.36 HREW 0.20 HREW 0.50 HREW	
	HRKW: Low Middle High	0.55 HRKW 0.54 HRKW 0.42 HRKW	
	HR15N: Low Middle High	0.44 HR15N 0.43 HR15N 0.53 HR15N	
	HR30N: Low Middle High	0.41 HR30N 0.30 HR30N 0.28 HR30N	
	HR45N: Low Middle High	0.51 HR45N 0.52 HR45N 0.25 HR45N	
	HR15TW: Low Middle High	0.30 HR15TW 0.27 HR15TW 0.47 HR15TW	

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Indirect Verification of Rockwell Hardness Testers & Rockwell Superficial Hardness <sup>3</sup> (cont)	HR30TW: Low Middle High  HR45TW: Low Middle High	0.59 HR30TW 0.38 HR30TW 0.29 HR30TW  0.72 HR45TW 0.47 HR45TW 0.40 HR45TW	Indirect verification method per ASTM E18
Partial Direct Verification of Rockwell Hardness Testers <sup>3</sup> –  Verification of the Test Force	(3 to 150) kgf	(0.047 + 0.41 %) kgf	ASTM E18, ISO 6508  Verification of the test force is by load cell per the method of ASTM E4
Indirect Verification of Brinell Hardness Testers at Test Conditions <sup>3</sup>	HBW 10/500/15 (HBW 217)  HBW 10/1500/15 (HBW 186)  HBW 10/3000/15 (HBW 230)	5.0 HBW  1.9 HBW  2.8 HBW	Indirect verification method per ASTM E10
Direct Verification of Brinell Hardness Testers <sup>3</sup> –  Verification of the Test Force  Verification of the Device for Measuring Indentation Diameters	≤ 3000 kgf  (0 to 7) mm	(0.034 + 0.1 %) kgf  0.29 mm	Direct verification method per ASTM E10, E110, ISO 6506  Verification of the test force is by load cell per the method of ASTM E4  Stage micrometer

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Indirect Verification of Equotip (Leeb) Hardness Testers <sup>3</sup>	(780 to 792) Ld	16 Ld	ASTM A956
Indirect Verification of Microindentation Hardness Testers <sup>3</sup> –  Knoop  Vickers	(545 HK, 733 HK)  (531 HV, 729 HV)	(13 HK, 20 HK)  (20 HV, 25 HV)	Indirect verification method per ASTM E384
Partial Direct Verification of Microindentation Hardness Testers <sup>3</sup> – (Vickers & Knoop)  Verification of the Test Force  Verification of the Device for Measuring Indentation Diagonals	(50 to 1000) gf  (0 to 7) mm	(0.14 + 0.11 %) gf  0.29 mm	ASTM E384, ISO 6507  Verification of the test force is by load cell per the method of ASTM E4  Stage micrometer
Partial Direct Verification of Macroindentation Hardness Testers <sup>3</sup> – (Vickers)  Verification of the Test Force  Verification of the Device for Measuring Indentation Diagonals	(1 to 120) kg  (0 to 7) mm	(0.12 + 0.35 %) kg  0.29 mm	ASTM E384, ISO 6507  Verification of the test force is by load cell per the method of ASTM E4  Stage micrometer

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Torque Wrenches <sup>3</sup>	(5 to 250) lbf·in	(0.15 + 2.2 %) lbf·in	Torque calibrator
Pressure/Vacuum Gauges <sup>3</sup>	(-15 to 30) psig (30 to 500) psig	(0.0082 + 0.014 %) psi (0.15 + 0.0042 %) psi	Pressure calibrators
RPM <sup>3</sup> – Calibration of Tachometer	(1 to 10 000) RPM	(0.58 + 0.093 %) RPM	KMOON MHS 5200A
Calibration of Photohelic, Magnehelic, & Environmental Gauge <sup>3</sup>	(0 to 27) in·H <sub>2</sub> O	(0.16 + 0.001/in·H <sub>2</sub> O) in·H <sub>2</sub> O	Fluke 750 P00 & 750 P02 pressure modules
Foundry Sand Test Equipment <sup>3</sup> –			
Core Scratch Hardness Tester	Up to 0.1 in	0.0030 in	Hardness block flat surface test
Green Sand Hardness Tester	(0 to 100) B & C	0.56 units on B & C scale	Flat surface test for 100 on scale B&C
Permmeter	(0 to 500) perms	0.94 perms	Perm standard
Sand Rammer	(0 to 6) inch	0.0014 in	Caliper, impact rings
Sand Strength Tester	(0 to 500) psi	(1.5 + 0.48 %) psi	Certified load cell
Moisture Teller	(0 to 400) °F	1.2 °F	Temperature calibrator

IV. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
IR Temperature – Measure <sup>3</sup>	(125 to 900) °F	(1.3 + 0.002/°F) °F	Black body Type N thermocouple, Fluke 7526A
Temperature – Measure <sup>3</sup>	(200 to 600) °C (600 to 1372) °C	0.93 °C (-1.7 + 0.005/ °C) °C	Martel MC-1010 with thermocouples such as K-type
Liquid-in-Glass Thermometers & Temperature Measuring Devices <sup>3</sup>	(0 to 100) °C (0 to 100) °C	0.43 °C 0.063 °C	Fluke 525B, RTD sensor
Relative Humidity – Measure	(10 to 95) % RH	2.8 % RH	Rotronics HygroClip2
Relative Humidity – Generate	11 % RH 33 % RH 75 % RH	0.50 % RH 0.70 % RH 0.90 % RH	Salt solutions
Climatic – Freezers, Refrigerators, Incubators, Ovens, and Furnaces			
Type K	(0 to 1372) °C	0.75 °C	Martel MC1010
Type N	(0 to 1300) °C	0.58 °C	
Type T	(-250 to 400) °C	1.0 °C	
Thermal Processing Equipment – Temperature Uniformity Surveys (TUS)	(0 to 1372) °C	(1.7 + 0.42 %) °C	(AMS2750, AIAG CQI-9) Yokogawa DX/GX1012 K-Type TC



## V. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> ( $\pm$ )	Comments
Frequency – Measure <sup>3</sup>	(3.0 to 1000) Hz	(-0.0053 + 0.31 %) Hz	Fluke 8845A
Frequency – Measuring Equipment <sup>3</sup>	(1.0 to 1000) Hz	(0.0057 + 0.006 %) Hz	KMOON MHS-5200A signal generator
Time Interval – Measure	1 s to 24 h	(0.62 + 0.000 47 %) s	Traceable 1051 stopwatch

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC, percentages are to be read as percent of reading, unless otherwise noted.

<sup>5</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>6</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

<sup>7</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.



# Accredited Laboratory

A2LA has accredited

## AMERICAN INSTRUMENT CORPORATION

*Hartland, WI*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 7<sup>th</sup> day of December 2022

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1354.01  
Valid to December 31, 2024

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*